

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1939
CALIBRATION DATE: 01-Sep-10p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4511

Voffset = -0.5375

Tau20 = 1.38

A = -2.6981e-003

B = 1.3229e-004

C = -2.2253e-006

E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.30000e-2

D2 = -4.64803e-2 H2 = 5.00000e+3

H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.26	6.00	0.02	0.863	1.26	-0.00
1.26	2.00	0.01	0.829	1.27	0.00
1.27	12.00	0.02	0.918	1.27	0.00
1.27	20.00	0.02	0.989	1.27	-0.00
1.28	26.00	0.02	1.048	1.28	-0.00
1.28	30.00	0.02	1.088	1.28	0.00
4.18	20.00	0.02	2.023	4.18	-0.00
4.18	2.00	0.01	1.500	4.18	-0.00
4.19	12.00	0.02	1.789	4.18	-0.00
4.19	6.00	0.02	1.616	4.19	0.00
4.19	26.00	0.02	2.205	4.19	-0.00
4.19	30.00	0.02	2.333	4.19	0.00
6.73	30.00	0.02	3.422	6.73	-0.00
6.77	12.00	0.02	2.564	6.78	0.00
6.78	20.00	0.02	2.946	6.78	-0.00
6.79	26.00	0.02	3.245	6.80	0.00
6.79	6.00	0.02	2.288	6.80	0.00
6.83	2.00	0.01	2.110	6.83	-0.00

$$\text{Oxygen (ml/l)} = \text{Soc} * (\text{V} + \text{Voffset}) * (1.0 + \text{A} * \text{T} + \text{B} * \text{T}^2 + \text{C} * \text{T}^3) * \text{OxSol}(\text{T},\text{S}) * \exp(\text{E} * \text{P} / \text{K})$$

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)

